


JUN 24 2010

VERIFICATION OF CALIBRATION REPORT
of DataMaster edm Breath Test Instrument
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster edm S/N	130133✓
A	Name	BRYAN BARLOW	ID# 4548 Date 6/18/10✓
	Agency	Alaska State Troopers	Phone # 907-783-0972
	Instrument Location	Girdwood AST✓	
B	Alco S/N	X172966✓	Target Value .084✓ High Pressure 800
	Alco Test Value Average	.084✓ 1 st Alco	.085✓ 2 nd Alco
	Signature		
(OVER)			


BMB
6/24/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 23rd day of July, 2010.


(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

JUNE 18, 2010

OPERATOR'S NAME:

BARLOW BRYAN W

OPERATOR'S NUMBER: 4548

SUBJECT'S LAST NAME:

CALIBRATION

SUBJECT'S FIRST NAME/NI:

VERIFICATION OF

O.L. #: 1234567

DEPT/AGENCY: 60108

CASE/REPORT: 10-00000

TEST TYPE: U

ALCO TARGET VALUE: .084

ALCO S/N: X172966

--- BREATH ANALYSIS ---

.084 ADJUSTED FOR 30.51 in
ALCO TARGET .085 21:02
BLANK TEST .000 21:03
INTERNAL STANDARD VERIFIED 21:03
ALCO TV 30.51 in .084 21:03
BLANK TEST .000 21:04
SUBJECT SAMPLE .000 21:05
BLANK TEST .000 21:06
ALCO TV 30.51 in .085 21:06
BLANK TEST .000 21:07

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

JUNE 18, 2010

TIME 20:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 30.51 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""#\$%&'()*+,-./0123456789:;<=>?@ABCDEF6
HIJ KLMNOPQRSTUVWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■